# MODELLING OF THERMAL DAMAGE IN LASER IRRADIATED TISSUE WITH EMBEDDED NANOPARTICLES

#### M. Jasiński


Institute of Computational Mechanics and Engineering, Silesian University of Technology, Gliwice, Poland e-mail: marek.jasinski@polsl.pl

## 1. Formulation of the problem

The purpose of this paper is to analyze the phenomena occurring in the laser-treated soft tissue wherein the cloud of nanoparticles is placed. The 2D domain of homogeneous biological tissue subjected to the laser action (Fig.1) is considered. The analysis is based on the Pennes bioheat transfer equation in the form [1]

(1) 
$$\mathbf{x} \in \Omega: \quad c\dot{T} = \lambda \nabla^2 T + Q_{perf} + Q_{las} + Q_{met}$$

where  $\lambda$  [Wm<sup>-1</sup>K<sup>-1</sup>] is the thermal conductivity, c [Jm<sup>-3</sup>K<sup>-1</sup>] is the volumetric specific heat,  $Q_{perf}$ ,  $Q_{met}$  and  $Q_{las}$  [Wm<sup>-3</sup>] are the heat sources connected with the perfusion, metabolism and laser radiation, respectively, while  $T = T(\mathbf{x}, t)$  is the temperature. Equation (1) is supplemented by appropriate boundary conditions: Robin condition on the external tissue surface  $\Gamma_0$  and no-flux condition on the internal tissue surface  $\Gamma_c$ . The initial distribution of temperature is also known.



Figures 1: The domain considered.

In order to determine the internal heat source concerning information about laser irradiation the collimated and diffuse part of fluence rate must be determined. The diffuse fluence rate  $\phi_d$  is calculated on the base of the steady-state optical diffusion equation [2,3]

(2) 
$$\mathbf{x} \in \Omega: \quad D\nabla^2 \phi_d(\mathbf{x}) - \mu_a \phi_d(\mathbf{x}) + \mu'_s \phi_c(\mathbf{x}) = 0$$

while the collimated fluence rate  $\phi_c$  is given as [1]

$$\phi_c(\mathbf{x}) = \phi_0 \exp\left(-\frac{2x_2^2}{r^2}\right) \exp(-\mu_t' x_1)$$

where D [m] is the diffusion coefficient,  $\mu_a$ ,  $\mu'_s$  and  $\mu'_t$  [m<sup>-1</sup>] are the absorption, effective scattering and effective attenuation coefficient of tissue, respectively,  $\phi_0$  [Wm<sup>-2</sup>] is the surface irradiance of laser, r is the radius of laser beam.

The final form of the source function connected with the laser heating is described by the formula

(4) 
$$Q_{las}(\mathbf{x},t) = \mu_a \, \phi(\mathbf{x}) \, p(t)$$

where  $\phi(\mathbf{x})$  [Wm<sup>-2</sup>] is the sum of collimated and diffuse parts of fluence rate and and p(t) is the function equal to 1 when the laser is *on* and equal to 0 when the laser is *off*.

Session: Biomechanics Abstract P032

Damage of biological tissue resulting from temperature elevation is modelled by Arrhenius injury integral, defined as [2]

(5) 
$$\Psi(\mathbf{x}, t^F) = \int_{0}^{t^F} P \exp\left[-\frac{E}{RT(\mathbf{x}, t)}\right] dt$$

where R [J mole<sup>-1</sup>K<sup>-1</sup>] is the universal gas constant, E [J mole<sup>-1</sup>] is the activation energy and P [s<sup>-1</sup>] is the pre-exponential factor. The criterion for tissue necrosis is  $\Psi(\mathbf{x}) \ge 1$ .


## 2. Results of computations

As was mention previously the 2D domain of homogeneous biological tissue subjected to the laser action was considered. Two simulations were carried out – with and without the cloud of nanoparticles which was situated near the external surface of the tissue (Fig. 1). It should be pointed out that optical properties of tissue with nanoparticles were calculated on the basis of formulas [2,3]

(6) 
$$\mu_{ap} = \mu_a + 0.75 f_v \frac{Q_a}{a}, \quad \mu_{sp} = \mu_s + 0.75 f_v \frac{Q_s}{a}$$

where  $Q_a$  and  $Q_s$  are the dimensionless efficiency factor of absorption and scattering for single particles, respectively,  $f_v$  is the volume fraction of nanoparticles while a is the particle radius.

The bioheat problem (1) has been solved using the 1st scheme of the BEM for 2D transient heat diffusion while the optical diffusion equation (2) has been solved by the finite difference method.



Figures 2: Distribution of diffuse fluence rate  $\phi_d$ , temperature and Arrhenius integral after 10 s.

**Acknowledgments** The research is funded from the projects Silesian University of Technology, Faculty of Mechanical Engineering.

#### References

- [1] F. Fanjul-Vélez, O.G. Romanov, J.L. Arce-Diego. Efficient 3D numerical approach for temperature prediction in laser irradiated biological tissues. *Comput. Biol. Med.*, 39:810, 2009.
- [2] L.A. Dombrovsky, V. Timchenko, M. Jackson. Indirect heating strategy for laser induced hyperthermia: An advanced thermal model. *Int. J. Heat Mass Transfer*, 55: 4688, 2012.
- [3] Y. Bayazitoglu, S. Kheradmand, T.K. Tullius. An overview of nanoparticle assisted laser therapy. *Int. J. Heat Mass Transfer*, 67: 469, 2013.